



# Technická inšpekcia, a.s.

Trnavská cesta 56, 821 01 Bratislava, Slovenská republika

## CERTIFICATE No.: 3916/1/2023

to technical file of equipment issued according of EN ISO/IEC 17020

Date of issue: 8<sup>th</sup> December 2023

Name of equipment:

1. BENCH TOP LABORATORY CENTRIFUGE: T-8M, SP-8M, T-8BL, SP-8BL, T-9BL
2. REFRIGERATED CENTRIFUGE MACHINE: SP-50, SP-60, SP-70BL, SP80BL, T-50, T-60, T-70BL
3. CYTO CENTRIFUGE: CYTO-8
4. BLOOD BAG CENTRIFUGE: BBC-6
5. BLOOD BAG TUBE SEALER: BBS-10
6. BOD INCUBATOR: BOD-108
7. LAMINAR AIR FLOW: AM-4
8. ELECTRIC SYRINGE AND NEEDLE DESTROYER: ND-100, ND-120

Technical characteristics 220-230 V AC, 15-20 A, 50 Hz

Manufacturer: LABY INSTRUMENTS INDUSTRY

Address: 62, INDUSTRIAL ESTATE, AMBALA CANTT-133006, HARYANA, INDIA

Applicant: See manufacturer

Basis for certificate issuing: Inspection report No. 3916/1/2023

Applicable standards: EN 61010-1:2010 + A1:2019

After evaluation of submitted documentation of equipment (TCF) in accordance with Directive 2014/35/EU, Technická inšpekcia, a. s. issues this  
certificate

the technical file of equipment fulfills safety requirements and it is in accordance with Low Voltage Directive 2014/35/EU.

The manufacturer may indicate compliance to this directive by signing EU declaration of conformity himself according Article 15 of this directive and applying the CE marking  to the equipment according to Article 16 and 17 of this directive.

The manufacturer will inform Technická inšpekcia, a.s. about every modification that can affect safety of the equipment.

This certificate is valid to 8<sup>th</sup> December 2026

Bratislava, 8<sup>th</sup> December 2023

On behalf of Technická inšpekcia, a.s.



Ing. Jozef Kamzík  
Director of Regional office Bratislava

CERT8-056



A 015295